
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002502
12:40 10/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:28
PRESSURIZING 2:31
STERILIZING 3:35
136.4 C 227kPa 3:35
136.4 C 227kPa 4:05
136.4 C 227kPa 4:35
136.2 C 225kPa 5:05
136.4 C 227kPa 5:35
136.3 C 225kPa 6:05
136.2 C 223kPa 6:35
136.3 C 226kPa 7:05
136.3 C 226kPa 7:05
VENTING 7:05
AIR DRYING 7:59
CYCLE COMPLETE 9:55

Digital Signature #
236C8EEED7D167CD
