
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002385
11:05 16/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:48
STERILIZING 3:48
136.4 C 224kPa 3:48
136.2 C 225kPa 4:18
136.3 C 225kPa 4:48
136.4 C 221kPa 5:18
136.4 C 223kPa 5:48
136.4 C 226kPa 6:18
136.4 C 226kPa 6:48
136.5 C 225kPa 7:18
136.5 C 225kPa 7:18
VENTING 7:18
AIR DRYING 8:08
CYCLE COMPLETE 8:22

Digital Signature #
7F98FB29834B134D
