

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1903

Name: RITTER 1  
Sterilizer ID: V2178198  
05/03/2022 02:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.9 F 0.0  
2:00 170.6 F 2.7  
4:00 190.2 F 6.9  
6:00 204.8 F 12.0  
8:00 222.9 F 6.2  
10:00 235.4 F 11.5  
12:00 233.8 F 8.3  
14:00 241.6 F 10.8  
16:00 245.6 F 13.1  
18:00 258.5 F 20.1  
20:00 270.0 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.6 F 28.7  
2:00 272.5 F 29.2  
3:00 271.5 F 28.6  
4:00 272.7 F 28.9  
  
Min 271.5 F 28.4  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.6 F 0.0  
5:00 223.4 F 0.0  
10:00 222.7 F 0.0  
15:00 224.3 F 0.0  
20:00 236.4 F 0.0  
25:00 232.8 F 0.0  
30:00 215.5 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 20:51  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:20  
CYCLE COMPLETE

05/03/2022 03:07 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02