
Midmark M11 - v1.0.5
Total Cycles: 888

Name: RITTER2
Sterilizer ID: V2313228
03/21/2022 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	70.8 F	0.0
2:00	82.5 F	0.0
4:00	116.5 F	0.0
6:00	150.7 F	0.2
8:00	184.2 F	0.5
10:00	206.3 F	1.1
12:00	216.0 F	1.9
14:00	219.9 F	2.5
16:00	222.0 F	3.1
18:00	223.7 F	3.7
20:00	225.3 F	4.3
22:00	227.3 F	5.0
24:00	230.9 F	6.4
26:00	234.5 F	7.9
28:00	245.0 F	12.7
30:00	235.0 F	8.2
32:00	244.9 F	11.8
34:00	240.8 F	10.8
36:00	232.7 F	7.2
38:00	248.2 F	14.4

C533: HEATUP MODE
STEAM TEMP LOW
CHAMBER TEMP 261F
CHAMBER PRESS 21.5PSI
** ITEMS NOT STERILE **

03/21/2022 10:52 AM