
Midmark M11 - v1.0.0
Total Cycles: 642

Name: MIDMARK
Sterilizer ID: V2178198
01/28/2021 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 116.3 F 0.0
2:00 118.4 F 0.0
4:00 124.2 F 1.2
6:00 156.2 F 4.6
8:00 186.7 F 9.0
10:00 233.8 F 9.7
12:00 226.2 F 9.2
14:00 238.5 F 9.2
16:00 242.1 F 11.6
18:00 239.0 F 9.6
20:00 252.4 F 16.2
22:00 263.9 F 22.9
24:00 270.0 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 272.2 F 28.6
2:00 272.2 F 28.7
3:00 272.0 F 28.9
4:00 271.4 F 28.8

Min 271.3 F 28.6
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.3 F 0.0
5:00 237.0 F 0.0
10:00 220.7 F 0.0
15:00 213.1 F 0.0
20:00 200.1 F 0.0
25:00 232.2 F 0.0
30:00 234.7 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 24:30
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:02:40
CYCLE COMPLETE

01/28/2021 03:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____