
Midmark M11 - v1.0.0
Total Cycles: 4129

Name: RITTER 1
Sterilizer ID: V2178198
12/18/2023 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.1 F 0.0
2:00 160.6 F 2.0
4:00 173.2 F 5.8
6:00 189.0 F 10.1
8:00 208.6 F 5.0
10:00 220.5 F 7.6
12:00 232.0 F 12.2
14:00 232.8 F 8.6
16:00 240.8 F 11.2
18:00 244.3 F 12.9
20:00 256.6 F 19.3
22:00 267.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.4 F 29.3
2:00 271.6 F 28.9
3:00 272.4 F 29.3
4:00 271.4 F 28.7

Min 271.4 F 28.6
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.5 F 0.0
5:00 216.6 F 0.0
10:00 229.8 F 0.0
15:00 228.3 F 0.0
20:00 224.3 F 0.0
25:00 226.9 F 0.0
30:00 232.3 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 23:22
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:01:56
CYCLE COMPLETE

12/18/2023 11:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____