
Midmark M11 - v1.0.0
Total Cycles: 3347

Name: RITTER 1
Sterilizer ID: V2178198
06/07/2023 07:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 79.7 F 0.0
4:00 90.0 F 0.0
6:00 118.9 F 0.0
8:00 153.0 F 0.0
10:00 178.7 F 2.6
12:00 195.6 F 6.0
14:00 210.3 F 9.3
16:00 222.2 F 12.9
18:00 224.8 F 6.9
20:00 236.3 F 11.2
22:00 232.2 F 7.7
24:00 245.4 F 13.5
26:00 243.5 F 12.1
28:00 255.7 F 18.4
30:00 266.6 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.7 F 29.0
2:00 271.5 F 28.2
3:00 272.6 F 29.1
4:00 272.1 F 28.5

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 201.7 F 0.0
10:00 208.6 F 0.0
15:00 207.6 F 0.0
20:00 209.5 F 0.0
25:00 207.9 F 0.0
30:00 203.2 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:37
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:09:40
CYCLE COMPLETE

06/07/2023 08:45 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____