
Midmark M11 - v1.0.0
Total Cycles: 1489

Name: RITTER 1
Sterilizer ID: V2178198
12/29/2021 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.5 F 0.0
2:00 200.6 F 3.8
4:00 206.3 F 8.5
6:00 218.1 F 10.0
8:00 228.5 F 9.0
10:00 236.5 F 10.2
12:00 241.3 F 12.1
14:00 240.4 F 10.6
16:00 253.6 F 17.3
18:00 264.8 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.1
2:00 271.8 F 28.9
3:00 272.4 F 29.0
4:00 271.7 F 28.9

Min 271.5 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 226.3 F 0.0
10:00 238.1 F 0.0
15:00 230.6 F 0.0
20:00 228.4 F 0.0
25:00 229.8 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:59
HEATING: 19:51
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:01
CYCLE COMPLETE

12/29/2021 03:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____