
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002073
15:04 19/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:37
STERILIZING 4:40
136.4 C 225kPa 4:40
136.4 C 224kPa 5:10
136.4 C 220kPa 5:40
136.3 C 225kPa 6:10
136.1 C 222kPa 6:40
136.3 C 223kPa 7:10
136.2 C 221kPa 7:40
136.2 C 223kPa 8:10
136.2 C 223kPa 8:10
VENTING 8:10
AIR DRYING 9:04
CYCLE COMPLETE 11:42

Digital Signature #
1B6C4B7DB734A227
