

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4344

Name: RITTER2  
Sterilizer ID: V2313228  
02/10/2025 12:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.1 F 0.0  
2:00 149.9 F 1.2  
4:00 151.0 F 3.4  
6:00 181.9 F 6.7  
8:00 192.9 F 10.6  
10:00 224.0 F 6.6  
12:00 223.0 F 7.4  
14:00 232.2 F 11.5  
16:00 232.1 F 7.7  
18:00 243.6 F 12.8  
20:00 240.4 F 10.7  
22:00 251.1 F 16.1  
24:00 260.3 F 21.4  
26:00 269.0 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 271.9 F 29.7  
2:00 271.5 F 29.1  
3:00 272.6 F 29.6  
4:00 271.4 F 29.6

Min 270.5 F 28.6  
Max 273.2 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
5:00 224.4 F 0.0  
10:00 209.6 F 0.0  
15:00 218.5 F 0.0  
20:00 215.3 F 0.0  
25:00 214.2 F 0.0  
30:00 200.3 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 27:06  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:06:03  
CYCLE COMPLETE

02/10/2025 01:07 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_