
Midmark M11 - v1.0.0
Total Cycles: 3444

Name: RITTER 1
Sterilizer ID: V2178198
06/26/2023 11:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.7 F 0.0
2:00 201.3 F 2.9
4:00 205.2 F 7.8
6:00 222.4 F 13.5
8:00 230.3 F 8.9
10:00 239.4 F 10.5
12:00 242.7 F 12.3
14:00 241.5 F 11.2
16:00 254.8 F 17.9
18:00 266.2 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.7 F 29.0
2:00 271.3 F 28.3
3:00 272.7 F 29.2
4:00 272.0 F 28.6

Min 271.3 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 229.6 F 0.0
10:00 224.8 F 0.0
15:00 233.5 F 0.0
20:00 229.1 F 0.0
25:00 219.4 F 0.0
30:00 211.8 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 19:37
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

06/26/2023 12:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____