

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 006594  
10:52 13/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:35  
PRESSURIZING 2:39  
STERILIZING 3:48  
136.5 C 224kPa 3:48  
136.4 C 220kPa 4:18  
136.4 C 223kPa 4:48  
136.4 C 224kPa 5:18  
136.4 C 225kPa 5:48  
136.3 C 223kPa 6:18  
136.4 C 221kPa 6:48  
136.0 C 221kPa 7:18  
136.0 C 221kPa 7:18  
VENTING 7:18  
AIR DRYING 8:10  
CYCLE COMPLETE 8:18  
-----

Digital Signature #  
4BED6BF26AB8ED9A  
-----