

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000184  
11:47 06/07/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:38  
PRESSURIZING 4:38  
STERILIZING 6:02  
136.4 C 223kPa 6:02  
136.2 C 223kPa 6:32  
136.2 C 222kPa 7:02  
136.3 C 224kPa 7:32  
136.4 C 225kPa 8:02  
136.2 C 220kPa 8:32  
136.4 C 225kPa 9:02  
136.3 C 224kPa 9:32  
136.3 C 224kPa 9:32  
VENTING 9:32  
AIR DRYING 10:34  
CYCLE COMPLETE 15:39

-----  
Digital Signature #  
CF68E69E70CA7485  
-----