
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006753
10:22 18/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:07
STERILIZING 4:19
136.4 C 220kPa 4:19
136.4 C 224kPa 4:49
136.4 C 219kPa 5:19
136.4 C 223kPa 5:49
136.4 C 220kPa 6:19
136.2 C 223kPa 6:49
136.4 C 220kPa 7:19
136.0 C 220kPa 7:49
136.0 C 220kPa 7:49
VENTING 7:49
AIR DRYING 8:41
CYCLE COMPLETE 10:55

Digital Signature #
0E01690156F151B5
