
Midmark M11 - v1.0.0
Total Cycles: 1682

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2022 04:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.5 F 0.0
2:00 196.5 F 3.8
4:00 209.2 F 8.7
6:00 224.2 F 7.7
8:00 232.2 F 9.4
10:00 239.4 F 9.1
12:00 244.5 F 12.9
14:00 243.5 F 11.8
16:00 256.8 F 18.7
18:00 268.3 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 28.9
2:00 272.0 F 28.3
3:00 271.7 F 28.5
4:00 272.6 F 28.9

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 221.7 F 0.0
10:00 224.6 F 0.0
15:00 230.2 F 0.0
20:00 228.8 F 0.0
25:00 227.3 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 2:07
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

03/01/2022 08:13 AM
ACCEPTED: X
REJECTED:
Operator: U02