
Midmark M11 - v1.0.5
Total Cycles: 2267

Name: RITTER2
Sterilizer ID: V2313228
04/19/2023 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.7 F 0.0
2:00 193.7 F 3.7
4:00 211.8 F 9.4
6:00 224.7 F 7.7
8:00 236.4 F 11.8
10:00 237.7 F 10.2
12:00 237.5 F 9.5
14:00 252.5 F 16.9
16:00 265.0 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.1 F 29.0
2:00 272.5 F 29.5
3:00 271.6 F 28.7
4:00 272.5 F 29.3

Min 271.5 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 229.3 F 0.0
10:00 238.8 F 0.0
15:00 229.0 F 0.0
20:00 225.6 F 0.0
25:00 224.5 F 0.0
30:00 223.0 F 0.0

Drying Complete

FILLING: 2:33
HEATING: 17:43
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:27
CYCLE COMPLETE

04/19/2023 11:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____