
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003530
12:11 26/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:21
STERILIZING 4:18
136.0 C 220kPa 4:18
136.4 C 225kPa 4:48
136.4 C 224kPa 5:18
136.3 C 223kPa 5:48
136.2 C 223kPa 6:18
136.4 C 221kPa 6:48
136.2 C 222kPa 7:18
136.4 C 224kPa 7:48
136.4 C 224kPa 7:48
VENTING 7:48
AIR DRYING 8:45
CYCLE COMPLETE 8:47

Digital Signature #
F41AFA9858748B3B
