
Midmark M11 - v1.0.5
Total Cycles: 916

Name: RITTER2
Sterilizer ID: V2313228
03/28/2022 02:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.8 F 0.0
2:00 151.5 F 0.0
4:00 179.7 F 0.7
6:00 206.5 F 4.9
8:00 221.2 F 10.0
10:00 228.2 F 8.2
12:00 238.1 F 12.0
14:00 238.7 F 10.5
16:00 238.2 F 9.8
18:00 252.8 F 17.1
20:00 265.0 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.8 F 29.6
2:00 272.0 F 28.9
3:00 271.5 F 29.0
4:00 272.8 F 29.7

Min 271.3 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 230.3 F 0.0
10:00 204.5 F 0.0
15:00 218.5 F 0.0
20:00 216.7 F 0.0
25:00 223.6 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:50
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:00:09
CYCLE COMPLETE

03/28/2022 03:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____