

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3996

Name: RITTER 1  
Sterilizer ID: V2178198  
11/08/2023 12:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.2 F 0.0  
2:00 160.1 F 2.1  
4:00 161.8 F 5.5  
6:00 183.1 F 9.4  
8:00 202.7 F 12.6  
10:00 214.3 F 6.0  
12:00 226.3 F 10.2  
14:00 232.7 F 8.1  
16:00 241.2 F 11.8  
18:00 238.9 F 10.0  
20:00 251.4 F 16.0  
22:00 262.2 F 22.2  
24:00 272.3 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.4 F 29.0  
2:00 271.7 F 28.5  
3:00 271.2 F 28.4  
4:00 272.2 F 29.1

Min 271.1 F 28.2  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
5:00 220.7 F 0.0  
10:00 230.2 F 0.0  
15:00 224.5 F 0.0  
20:00 214.2 F 0.0  
25:00 212.0 F 0.0  
30:00 209.1 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 24:32  
STERILIZING: 4:00  
VENTING: 3:46  
DRYING: 30:00  
TOTAL CYCLE: 01:03:24  
CYCLE COMPLETE

11/08/2023 01:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_