
Midmark M11 - v1.0.0
Total Cycles: 2382

Name: RITTER 1
Sterilizer ID: V2178198
09/29/2022 11:22 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.4 F 0.0
2:00 182.4 F 2.2
4:00 186.9 F 6.5
6:00 208.1 F 12.0
8:00 224.8 F 6.9
10:00 240.6 F 13.2
12:00 241.2 F 11.0
14:00 238.9 F 9.5
16:00 254.2 F 17.0
18:00 266.6 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.7
1:00 271.7 F 27.7
2:00 272.4 F 28.1
3:00 272.8 F 28.5

Min 271.4 F 27.5
Max 272.8 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 226.9 F 0.0
5:00 228.4 F 0.0
10:00 225.5 F 0.0
15:00 224.3 F 0.0
20:00 223.0 F 0.0
25:00 221.5 F 0.0
30:00 218.6 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 19:21
STERILIZING: 3:00
VENTING: 2:45
DRYING: 30:00
TOTAL CYCLE: 00:56:58
CYCLE COMPLETE

09/29/2022 12:18 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____