
Midmark M11 - v1.0.0
Total Cycles: 4918

Name: RITTER 1
Sterilizer ID: V2178198
09/18/2024 04:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.6 F 0.0
2:00 138.7 F 0.1
4:00 159.0 F 0.3
6:00 182.7 F 1.8
8:00 221.8 F 5.8
10:00 225.9 F 10.0
12:00 232.9 F 8.8
14:00 234.9 F 9.7
16:00 237.3 F 9.4
18:00 244.1 F 12.4
20:00 242.0 F 11.0
22:00 254.2 F 17.3
24:00 265.1 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.1 F 29.0
2:00 271.8 F 28.3
3:00 272.6 F 28.9
4:00 271.6 F 28.6

Min 270.8 F 27.9
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 234.5 F 0.0
10:00 229.2 F 0.0
15:00 229.3 F 0.0
20:00 225.9 F 0.0
25:00 228.1 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:50
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:04:35
CYCLE COMPLETE

09/18/2024 05:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____