

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2955

Name: RITTER 1  
Sterilizer ID: V2178198  
02/28/2023 02:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.8 F 0.0  
2:00 182.4 F 3.6  
4:00 203.0 F 8.3  
6:00 223.3 F 9.9  
8:00 232.0 F 9.8  
10:00 237.9 F 8.4  
12:00 244.3 F 12.6  
14:00 244.8 F 12.9  
16:00 258.2 F 20.0  
18:00 270.0 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.4 F 29.1  
2:00 271.3 F 28.5  
3:00 272.5 F 29.3  
4:00 271.7 F 28.6  
  
Min 271.2 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 218.9 F 0.0  
10:00 237.3 F 0.0  
15:00 238.4 F 0.0  
20:00 224.9 F 0.0  
25:00 219.4 F 0.0  
30:00 215.0 F 0.0  
Drying Complete

FILLING: 1:42  
HEATING: 18:53  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:57:50  
CYCLE COMPLETE

02/28/2023 03:11 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_