
Midmark M11 - v1.0.0
Total Cycles: 4521

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2024 09:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.2 F 0.0
2:00 163.3 F 2.3
4:00 173.5 F 5.6
6:00 193.0 F 9.5
8:00 212.4 F 8.1
10:00 219.5 F 6.9
12:00 230.9 F 11.3
14:00 231.8 F 7.4
16:00 243.1 F 13.0
18:00 242.4 F 11.4
20:00 254.6 F 17.5
22:00 265.1 F 23.8
24:00 270.9 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.8 F 28.8
2:00 272.1 F 29.0
3:00 272.3 F 28.6
4:00 272.1 F 29.1

Min 271.3 F 28.4
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 235.1 F 0.0
10:00 212.5 F 0.0
15:00 226.1 F 0.0
20:00 213.0 F 0.0
25:00 210.7 F 0.0
30:00 206.1 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 24:42
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:36
CYCLE COMPLETE

05/03/2024 10:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____