
Midmark M11 - v1.0.5
Total Cycles: 1243

Name: RITTER2
Sterilizer ID: V2313228
06/29/2022 02:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.0 F 0.0
2:00 155.7 F 3.0
4:00 170.0 F 6.7
6:00 193.4 F 11.5
8:00 213.5 F 5.1
10:00 227.8 F 10.2
12:00 232.7 F 8.0
14:00 242.2 F 13.2
16:00 242.3 F 12.0
18:00 255.1 F 18.7
20:00 266.4 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.4 F 29.8
2:00 272.0 F 29.2
3:00 271.4 F 29.2
4:00 272.8 F 29.8

Min 271.2 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 207.2 F 0.0
10:00 221.1 F 0.0
15:00 218.3 F 0.0
20:00 214.7 F 0.0
25:00 216.0 F 0.0
30:00 213.1 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:38
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

06/29/2022 03:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____