
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003899
9:56 28/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 3:01
STERILIZING 4:17
136.5 C 223kPa 4:17
136.3 C 224kPa 4:47
136.3 C 224kPa 5:17
136.1 C 218kPa 5:47
136.4 C 223kPa 6:17
136.2 C 222kPa 6:47
136.4 C 225kPa 7:17
136.4 C 222kPa 7:47
136.4 C 222kPa 7:47
VENTING 7:47
AIR DRYING 8:38
CYCLE COMPLETE 8:42

Digital Signature #
205E599E5A7CDBC9
