
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002197
12:29 01/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:08
STERILIZING 3:06
136.0 C 221kPa 3:06
136.4 C 224kPa 3:36
136.4 C 226kPa 4:06
136.3 C 225kPa 4:36
136.3 C 225kPa 5:06
136.3 C 226kPa 5:36
136.4 C 225kPa 6:06
136.3 C 225kPa 6:36
136.3 C 225kPa 6:36
VENTING 6:36
AIR DRYING 7:25
CYCLE COMPLETE 37:26

Digital Signature #
C56A9DFE1B8C6C22
