
Midmark M11 - v1.0.0
Total Cycles: 4869

Name: RITTER 1
Sterilizer ID: V2178198
09/05/2024 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.2 F 0.0
2:00 158.7 F 0.1
4:00 159.2 F 0.4
6:00 197.3 F 1.0
8:00 211.1 F 4.5
10:00 218.0 F 8.6
12:00 229.5 F 12.9
14:00 231.0 F 8.4
16:00 243.7 F 13.1
18:00 242.2 F 11.4
20:00 239.7 F 10.0
22:00 253.2 F 16.6
24:00 264.4 F 23.2
26:00 271.2 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 27.9
1:00 272.4 F 28.7
2:00 271.7 F 28.0
3:00 271.5 F 28.2
4:00 272.2 F 28.6

Min 271.1 F 27.8
Max 272.4 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 214.7 F 0.0
10:00 212.0 F 0.0
15:00 242.8 F 0.0
20:00 240.4 F 0.0
25:00 231.1 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 26:04
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:04:30
CYCLE COMPLETE

09/05/2024 11:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____