
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001551
12:59 30/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:09
PRESSURIZING 2:13
STERILIZING 3:11
136.0 C 220kPa 3:11
136.3 C 224kPa 3:41
136.4 C 224kPa 4:11
136.1 C 222kPa 4:41
136.4 C 225kPa 5:11
136.2 C 223kPa 5:41
136.4 C 224kPa 6:11
136.2 C 225kPa 6:41
136.2 C 225kPa 6:41
VENTING 6:41
AIR DRYING 7:31
CYCLE COMPLETE 37:31

Digital Signature #
7575E1AB91A5E01B
