
Midmark M11 - v1.0.0
Total Cycles: 3873

Name: RITTER 1
Sterilizer ID: V2178198
10/05/2023 01:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.4 F 0.0
2:00 176.2 F 3.5
4:00 189.5 F 7.6
6:00 218.9 F 12.8
8:00 226.7 F 7.2
10:00 239.0 F 12.2
12:00 236.1 F 9.4
14:00 238.3 F 9.5
16:00 248.0 F 14.1
18:00 259.7 F 20.5
20:00 270.5 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 272.2 F 28.6
2:00 272.4 F 28.9
3:00 272.7 F 29.2
4:00 271.8 F 29.0

Min 271.5 F 28.5
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.6 F 0.0
5:00 228.6 F 0.0
10:00 226.2 F 0.0
15:00 189.1 F 0.0
20:00 180.0 F 0.0
25:00 176.5 F 0.0
30:00 173.2 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 21:23
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:00:56
CYCLE COMPLETE

10/05/2023 02:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____