
Midmark M11 - v1.0.5
Total Cycles: 390

Name: RITTER2
Sterilizer ID: V2313228
09/23/2021 11:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.2 F 0.0
2:00 181.3 F 3.7
4:00 195.4 F 8.6
6:00 215.9 F 8.8
8:00 230.7 F 9.9
10:00 233.7 F 8.6
12:00 242.8 F 12.2
14:00 245.5 F 13.3
16:00 258.1 F 20.2
18:00 269.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 271.6 F 29.0
2:00 271.8 F 29.2
3:00 271.9 F 29.4
4:00 271.9 F 29.8

Min 271.0 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.1 F 0.0
5:00 230.9 F 0.0
10:00 223.8 F 0.0
15:00 218.5 F 0.0
20:00 215.3 F 0.0
25:00 212.5 F 0.0
30:00 207.0 F 0.0

Drying Complete

FILLING: 2:23
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

09/23/2021 12:08 PM
ACCEPTED: X
REJECTED:
Operator: U02