
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003574
13:12 02/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:46
STERILIZING 3:46
136.4 C 223kPa 3:46
136.2 C 224kPa 4:16
136.4 C 223kPa 4:46
136.4 C 222kPa 5:16
136.4 C 224kPa 5:46
136.4 C 223kPa 6:16
136.2 C 223kPa 6:46
136.4 C 223kPa 7:16
136.4 C 223kPa 7:16
VENTING 7:17
AIR DRYING 8:08
CYCLE COMPLETE 8:10

Digital Signature #
D3DA2DFCAD70EA26
