
Midmark M11 - v1.0.0
Total Cycles: 736

Name: RITTER 1
Sterilizer ID: V2178198
03/22/2021 10:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.1 F 0.0
2:00 183.6 F 2.5
4:00 191.4 F 6.9
6:00 211.2 F 13.0
8:00 226.7 F 8.3
10:00 239.9 F 11.1
12:00 240.8 F 12.1
14:00 240.3 F 10.8
16:00 254.0 F 17.8
18:00 265.7 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.8
1:00 273.0 F 29.4
2:00 272.1 F 29.0
3:00 272.9 F 29.4
4:00 271.9 F 28.8

Min 271.9 F 28.7
Max 273.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.0 F 0.0
5:00 236.2 F 0.0
10:00 230.4 F 0.0
15:00 220.5 F 0.0
20:00 218.3 F 0.0
25:00 219.9 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 19:38
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:27
CYCLE COMPLETE

03/22/2021 11:12 AM
ACCEPTED: X
REJECTED:
Operator: U01