
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000231
11:38 28/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:21
PRESSURIZING 3:24
STERILIZING 4:54
136.0 C 222kPa 4:54
136.2 C 232kPa 5:24
136.3 C 232kPa 5:54
136.4 C 233kPa 6:24
136.4 C 233kPa 6:54
136.4 C 231kPa 7:24
136.3 C 232kPa 7:54
136.4 C 233kPa 8:24
136.4 C 233kPa 8:24
VENTING 8:24
AIR DRYING 9:28
CYCLE COMPLETE 10:38

Digital Signature #
6010B4C78605E115
