
Midmark M11 - v1.0.5
Total Cycles: 1488

Name: RITTER2
Sterilizer ID: V2313228
09/14/2022 09:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 151.3 F 0.0
2:00 152.4 F 1.7
4:00 182.9 F 4.6
6:00 203.6 F 8.6
8:00 216.3 F 13.2
10:00 219.4 F 7.6
12:00 232.0 F 12.6
14:00 233.0 F 9.4
16:00 237.6 F 9.8
18:00 245.6 F 13.8
20:00 257.6 F 20.3
22:00 268.4 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.7 F 30.1
2:00 272.0 F 29.4
3:00 271.4 F 29.5
4:00 272.6 F 30.1

Min 271.2 F 29.2
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 215.6 F 0.0
10:00 225.7 F 0.0
15:00 224.7 F 0.0
20:00 223.0 F 0.0
25:00 221.9 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:17
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:24
CYCLE COMPLETE

09/14/2022 10:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____