
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001462
10:36 09/30/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:30
PRESSURIZING 6:09
STERILIZING 7:40
136.0 C 220kPa 7:40
136.5 C 223kPa 8:10
136.5 C 223kPa 8:40
136.4 C 222kPa 9:10
136.5 C 221kPa 9:40
136.4 C 223kPa 10:10
136.5 C 223kPa 10:40
136.4 C 223kPa 11:10
VENTING 11:10
AIR DRYING 12:14
CYCLE COMPLETE 42:15

Digital Signature #
BED40CE710AE198F
