

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000763  
8:35 28/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:20  
PRESSURIZING 3:23  
STERILIZING 4:33  
136.0 C 228kPa 4:33  
136.4 C 232kPa 5:03  
136.2 C 231kPa 5:33  
136.4 C 233kPa 6:03  
136.3 C 232kPa 6:33  
136.2 C 231kPa 7:03  
136.4 C 233kPa 7:33  
136.1 C 230kPa 8:03  
136.1 C 230kPa 8:03  
VENTING 8:03  
AIR DRYING 8:58  
CYCLE COMPLETE 9:00  
-----

Digital Signature #  
14E3BB80607499EE  
-----