

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 008522  
14:21 04/04/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:18  
PRESSURIZING 2:22  
STERILIZING 3:22  
136.4 C 222kPa 3:22  
135.8 C 220kPa 3:52  
136.3 C 222kPa 4:22  
136.4 C 224kPa 4:52  
136.4 C 222kPa 5:22  
136.0 C 221kPa 5:52  
136.2 C 221kPa 6:22  
136.4 C 223kPa 6:52  
136.4 C 223kPa 6:52  
VENTING 6:53  
AIR DRYING 7:44  
CYCLE COMPLETE 37:44

-----  
Digital Signature #  
57173741EAE3AB25  
-----