
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007909
7:50 02/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:21
PRESSURIZING 3:24
STERILIZING 4:28
136.0 C 220kPa 4:28
135.9 C 220kPa 4:59
136.4 C 220kPa 5:29
136.2 C 221kPa 5:59
136.4 C 223kPa 6:29
136.4 C 222kPa 6:59
136.3 C 221kPa 7:29
136.4 C 221kPa 7:59
136.4 C 221kPa 7:59
VENTING 7:59
AIR DRYING 8:50
CYCLE COMPLETE 38:51

Digital Signature #
B1DC4F726F6611BD
