
Midmark M11 - v1.0.0
Total Cycles: 4189

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2024 01:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.7 F 0.0
2:00 126.7 F 1.5
4:00 145.1 F 4.3
6:00 181.5 F 7.6
8:00 196.2 F 12.2
10:00 213.1 F 4.8
12:00 223.1 F 8.8
14:00 234.8 F 13.2
16:00 235.0 F 9.8
18:00 238.4 F 9.5
20:00 246.0 F 13.8
22:00 257.5 F 20.0
24:00 268.1 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.6 F 29.6
2:00 271.4 F 28.8
3:00 272.5 F 29.7
4:00 272.0 F 29.2

Min 271.2 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 211.2 F 0.0
10:00 216.7 F 0.0
15:00 217.4 F 0.0
20:00 216.6 F 0.0
25:00 211.4 F 0.0
30:00 208.3 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 25:19
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:04:22
CYCLE COMPLETE

01/10/2024 02:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____