

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4263

Name: RITTER 1  
Sterilizer ID: V2178198  
02/02/2024 11:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.1 F 0.0  
2:00 139.6 F 2.0  
4:00 156.6 F 4.9  
6:00 194.1 F 9.0  
8:00 209.5 F 11.7  
10:00 223.5 F 6.1  
12:00 231.9 F 10.6  
14:00 232.0 F 7.5  
16:00 242.3 F 11.9  
18:00 239.1 F 9.9  
20:00 251.7 F 16.0  
22:00 262.4 F 22.1  
24:00 272.4 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.0 F 28.4  
2:00 272.5 F 28.7  
3:00 272.5 F 28.9  
4:00 272.3 F 29.1

Min 271.5 F 28.4  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.2 F 0.0  
5:00 231.8 F 0.0  
10:00 213.7 F 0.0  
15:00 188.4 F 0.0  
20:00 182.5 F 0.0  
25:00 181.6 F 0.0  
30:00 180.7 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 25:10  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:03:51  
CYCLE COMPLETE

02/02/2024 12:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_