
Midmark M11 - v1.0.0
Total Cycles: 768

Name: RITTER 1
Sterilizer ID: V2178198
04/06/2021 07:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 72.4 F 0.0
4:00 80.1 F 0.0
6:00 103.0 F 0.1
8:00 133.9 F 0.2
10:00 172.1 F 0.3
12:00 199.8 F 4.0
14:00 211.2 F 7.0
16:00 222.8 F 10.5
18:00 232.0 F 8.3
20:00 237.8 F 10.8
22:00 233.7 F 7.8
24:00 242.3 F 11.0
26:00 246.5 F 13.3
28:00 259.1 F 20.1
30:00 270.1 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.0 F 28.5
2:00 271.8 F 28.7
3:00 271.6 F 28.1
4:00 272.3 F 28.9

Min 271.2 F 28.0
Max 272.3 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 202.9 F 0.0
10:00 220.1 F 0.0
15:00 212.0 F 0.0
20:00 204.7 F 0.0
25:00 242.1 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:51
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:08:42
CYCLE COMPLETE

04/06/2021 08:49 AM
ACCEPTED: X
REJECTED:
Operator: U01