
Midmark M11 - v1.0.5
Total Cycles: 149

Name: RITTER2
Sterilizer ID: V2313228
06/28/2021 12:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.0 F 0.0
2:00 140.1 F 2.0
4:00 156.7 F 5.5
6:00 184.2 F 10.0
8:00 203.6 F 5.7
10:00 219.4 F 8.0
12:00 232.7 F 13.0
14:00 236.4 F 10.4
16:00 236.6 F 8.8
18:00 249.0 F 15.3
20:00 260.5 F 21.8
22:00 270.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.3 F 29.5
2:00 272.6 F 29.9
3:00 271.8 F 29.2
4:00 272.4 F 29.7

Min 271.7 F 29.1
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.0 F 0.0
5:00 228.1 F 0.0
10:00 226.1 F 0.0
15:00 222.2 F 0.0
20:00 220.3 F 0.0
25:00 198.5 F 0.0
Dry Cycle Aborted

FILLING: 1:25
HEATING: 22:43
STERILIZING: 4:00
VENTING: 3:09
DRYING: 25:47
TOTAL CYCLE: 00:57:04

06/28/2021 01:14 PM
ACCEPTED: X
REJECTED:
Operator: U01