

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5029

Name: RITTER 1  
Sterilizer ID: V2178198  
10/28/2024 10:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.8 F 0.0  
2:00 164.4 F 0.3  
4:00 165.6 F 1.1  
6:00 184.8 F 4.8  
8:00 203.3 F 8.6  
10:00 214.5 F 12.7  
12:00 222.4 F 6.5  
14:00 234.8 F 11.0  
16:00 233.6 F 7.0  
18:00 244.9 F 13.1  
20:00 242.7 F 11.4  
22:00 254.2 F 17.2  
24:00 264.4 F 23.2  
26:00 271.9 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 272.3 F 28.7  
2:00 272.2 F 28.3  
3:00 271.3 F 27.9  
4:00 271.7 F 28.4

Min 271.1 F 27.8  
Max 272.5 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.6 F 0.0  
5:00 223.1 F 0.0  
10:00 229.5 F 0.0  
15:00 228.0 F 0.0  
20:00 228.7 F 0.0  
25:00 228.5 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 26:11  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:04:50  
CYCLE COMPLETE

10/28/2024 11:36 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_