
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.7uS / 0.4ppm
CYCLE NUMBER 001437
11:58 09/16/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:59
PRESSURIZING 7:02
STERILIZING 8:44
136.0 C 220kPa 8:44
136.5 C 222kPa 9:15
135.9 C 218kPa 9:45
136.4 C 218kPa 10:15
136.5 C 222kPa 10:45
136.4 C 222kPa 11:15
136.4 C 221kPa 11:45
136.4 C 221kPa 12:14
VENTING 12:15
AIR DRYING 13:19
CYCLE COMPLETE 29:45

Digital Signature #
DF555F9FB71CA4D4
