

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1655

Name: RITTER 1  
Sterilizer ID: V2178198  
02/22/2022 12:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.9 F 0.0  
2:00 182.2 F 2.7  
4:00 198.3 F 7.6  
6:00 217.4 F 13.4  
8:00 229.4 F 8.4  
10:00 238.8 F 10.3  
12:00 244.4 F 13.1  
14:00 244.3 F 12.4  
16:00 257.8 F 19.6  
18:00 269.3 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 271.9 F 28.4  
2:00 272.3 F 28.8  
3:00 272.4 F 29.1  
4:00 271.5 F 28.4  
  
Min 271.3 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.9 F 0.0  
5:00 221.6 F 0.0  
10:00 214.8 F 0.0  
15:00 218.2 F 0.0  
20:00 220.4 F 0.0  
25:00 223.7 F 0.0  
30:00 228.4 F 0.0  
Drying Complete

FILLING: 1:49  
HEATING: 18:56  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:55  
CYCLE COMPLETE

02/22/2022 01:04 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_