
Midmark M11 - v1.0.5
Total Cycles: 1798

Name: RITTER2
Sterilizer ID: V2313228
12/02/2022 11:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.1 F 0.0
2:00 167.4 F 0.0
4:00 167.6 F 0.2
6:00 198.8 F 5.1
8:00 223.1 F 11.1
10:00 228.1 F 7.5
12:00 242.1 F 13.5
14:00 242.7 F 12.0
16:00 241.8 F 11.2
18:00 256.2 F 18.6
20:00 268.5 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.1 F 28.4
2:00 272.8 F 29.0
3:00 272.2 F 29.0
4:00 271.9 F 28.4

Min 271.5 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 221.7 F 0.0
5:00 229.2 F 0.0
10:00 240.3 F 0.0
Dry Cycle Aborted

FILLING: 1:08
HEATING: 21:03
STERILIZING: 4:00
VENTING: 2:46
DRYING: 10:54
TOTAL CYCLE: 00:39:52

12/02/2022 12:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____