
Midmark M11 - v1.0.5
Total Cycles: 3170

Name: RITTER2
Sterilizer ID: V2313228
11/29/2023 06:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 124.6 F 0.0
2:00 120.8 F 0.6
4:00 141.0 F 3.0
6:00 170.2 F 6.5
8:00 188.4 F 10.8
10:00 209.0 F 4.8
12:00 225.1 F 8.4
14:00 236.7 F 12.9
16:00 234.6 F 9.7
18:00 236.8 F 9.2
20:00 246.9 F 14.3
22:00 258.9 F 20.9
24:00 270.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.5 F 29.8
2:00 271.4 F 29.0
3:00 272.4 F 29.9
4:00 272.0 F 29.3

Min 271.2 F 28.9
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.2 F 0.0
5:00 229.1 F 0.0
10:00 193.4 F 0.0
15:00 192.2 F 0.0
20:00 190.5 F 0.0
25:00 188.7 F 0.0
30:00 187.8 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 24:51
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:03:19
CYCLE COMPLETE

11/29/2023 07:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____