

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3801

Name: RITTER2  
Sterilizer ID: V2313228  
07/19/2024 11:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 120.3 F 0.0  
2:00 125.8 F 1.7  
4:00 142.1 F 4.3  
6:00 169.0 F 7.9  
8:00 191.5 F 11.9  
10:00 212.3 F 5.4  
12:00 219.6 F 9.9  
14:00 229.4 F 9.0  
16:00 237.0 F 11.2  
18:00 235.9 F 9.3  
20:00 249.6 F 15.7  
22:00 260.6 F 22.0  
24:00 270.8 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.5 F 30.0  
2:00 271.7 F 29.3  
3:00 271.4 F 29.7  
4:00 272.5 F 29.9

Min 271.0 F 29.1  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.2 F 0.0  
5:00 204.6 F 0.0  
10:00 207.3 F 0.0  
15:00 209.1 F 0.0  
20:00 206.2 F 0.0  
25:00 203.7 F 0.0  
30:00 205.1 F 0.0  
Drying Complete

FILLING: 1:00  
HEATING: 24:48  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:03:26  
CYCLE COMPLETE

07/19/2024 12:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_