
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001557
12:02 01/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:31
STERILIZING 3:33
136.0 C 220kPa 3:33
135.9 C 222kPa 4:03
136.3 C 225kPa 4:33
136.4 C 224kPa 5:03
136.4 C 226kPa 5:33
136.4 C 225kPa 6:03
136.2 C 224kPa 6:33
136.3 C 223kPa 7:03
136.3 C 223kPa 7:03
VENTING 7:03
AIR DRYING 7:57
CYCLE COMPLETE 37:57

Digital Signature #
C8DF26C3E761AE36
