
Midmark M11 - v1.0.5
Total Cycles: 2747

Name: RITTER2
Sterilizer ID: V2313228
08/09/2023 04:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.9 F 0.0
2:00 176.4 F 2.8
4:00 185.1 F 6.7
6:00 203.9 F 11.2
8:00 218.9 F 5.8
10:00 229.6 F 10.5
12:00 232.2 F 8.1
14:00 243.4 F 13.5
16:00 243.5 F 12.7
18:00 255.5 F 18.9
20:00 266.2 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.1
1:00 272.4 F 29.9
2:00 271.9 F 29.3
3:00 271.1 F 29.3
4:00 272.5 F 30.3

Min 271.0 F 29.0
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 226.7 F 0.0
10:00 225.8 F 0.0
15:00 219.8 F 0.0
20:00 219.8 F 0.0
25:00 216.0 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:01:00
CYCLE COMPLETE

08/09/2023 05:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____